



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/882,280	06/15/2001	Todd A. Thompson	9345.17121-CON 2	8625
26308	7590	06/09/2005	EXAMINER	
RYAN KROMHOLZ & MANION, S.C. POST OFFICE BOX 26618 MILWAUKEE, WI 53226			SMITH, RUTH S	
			ART UNIT	PAPER NUMBER
			3737	

DATE MAILED: 06/09/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary	Application No. 09/882,280	Applicant(s) THOMPSON ET AL.	
	Examiner Ruth S. Smith	Art Unit 3737	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 21 March 2005.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1,3,4,8,9 and 26-34 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1,3,4,8,9 and 26-34 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date <u>10/18/04, 3/21/05</u> . | 6) <input type="checkbox"/> Other: _____ |

Claim Rejections - 35 USC § 101

35 U.S.C. 101 reads as follows:

Whoever invents or discovers any new and useful process, machine, manufacture, or composition of matter, or any new and useful improvement thereof, may obtain a patent therefor, subject to the conditions and requirements of this title.

Claims 26-30 are rejected under 35 U.S.C. 101 because the claimed invention is directed to non-statutory subject matter. Claim 26 defines the halter assembly as being worn on the patient and including components worn about the shoulders and back of the patient thereby including the patient as part of the claimed invention. The inclusion of a living being as part of the claimed invention renders the claim non-statutory.

Claim Objections

Claims 31-34 are objected to because of the following informalities: In claim 31, line 12, applicator is misspelled. Appropriate correction is required.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 26,30 are rejected under 35 U.S.C. 102(e) as being anticipated by Talish et al ('070). The claims are directly readable on Talish et al which discloses a system for applying ultrasound to the thoracic cavity of a patient comprising an electric signal generating machine 12, an ultrasound applicator 16, and an assembly for placing the applicator on the patient. As seen in figure 1, the assembly can include a halter worn about the back and shoulders. The electrical signal generating machine is battery powered in that the device 12 is disclosed as being the same structure as that used in US Patent No. 5,556,372 which discloses use of a battery 32. The halter assembly includes components that are worn about the back that leave the chest on opposing

sides of the applicator uncovered which would allow placement of another treatment device on the chest.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

Claims 1,3,4,9,27-28,31-32,34 are rejected under 35 U.S.C. 103(a) as being unpatentable over Talish et al. Talish et al disclose a system for applying ultrasound to the thoracic cavity of a patient comprising an electric signal generating machine 12, an ultrasound applicator 16, and an assembly for placing the applicator on the patient. The assembly includes a quick release mechanism as seen at the end of straps 20 in figure 1 and a quick release material as seen by the Velcro in figure 5. As seen in figure 1, the assembly can include a halter worn about the chest and shoulders. The electrical signal generating machine is battery powered in that the device 12 is disclosed as being the same structure as that used in US Patent No. 5,556,372 which discloses use of a battery 32. Talish et al disclose, in column 9, that various modifications can be made to the structural configuration of the placement module. The placement module includes components that are worn about the back that leave the chest on opposing sides of the

housing uncovered which would allow placement of another treatment device on the chest. Therefore, in the absence of any showing of criticality, the specific arrangement of the assembly to provide stabilization of the housing would have been an obvious design choice of known functional equivalents in the art, particularly in view of Talish et al disclosing that various modifications can be made to the structural configuration of the placement module. With respect to claim 4, in the absence of any showing of criticality, the specific type of material selected would have been obvious based upon suitability for intended use.

Claims 8,29,33 are rejected under 35 U.S.C. 103(a) as being unpatentable over Talish et al ('070) as applied to claims 1,26,31 above, and further in view of Peterson et al. Talish et al fails to specifically disclose the operating parameters of the ultrasound energy. Peterson et al is just one example of many which disclose the operating parameters of the therapeutic ultrasound as set forth in the claims. It would have been obvious to one skilled in the art to have further modified Talish et al such that the operating parameters are as taught by Peterson et al in that such are well known operating parameters for therapeutic ultrasound which will not cause harm to the patient.

Response to Arguments

Applicant's arguments filed 3/21/05 have been fully considered but they are not persuasive. The placement module includes components that are worn about the back that leave the chest on opposing sides of the housing uncovered which would allow placement of another treatment device on the chest. Talish et al may not disclose the use of another treatment device, however, the claims merely set forth that placement of another device is possible and such a limitation is clearly seen in Talish et al.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP


§ 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Ruth S. Smith whose telephone number is 571-272-4745. The examiner can normally be reached on M-F 7:30 AM-4:00 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Brian Casler can be reached on 571-272-4956. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



Ruth S. Smith
Primary Examiner
Art Unit 3737

RSS